


CALIBRATION CERTIFICATE

Customer : 0155600
Identification number : 101132002
Gauge application : Incoming inspection
Standard of thread : Pipe threads acc. to ISO 7/2:2000 (EN 10226:2005)
Type of gauge : Parallel thread ring gauge for Taper threads
Thread designation : R 2
Pitch : 2,3090 mm
Measuring method : Inspection with checking or setting gauge

Gauge nominal values

Major diameter min. : 59,6140 mm
Pitch diameter Nennwert : 58,1350 mm
Minor diameter max. : 56,6560 mm
Minor diameter min. : 56,6200 mm
Distance max. to gauge plane : 0,2309 mm
Distance min. to gauge plane : -0,2309 mm



Measuring values Pitch diameter with taper check plug gauge

Testing Device ID	Nominal value [mm]	Actual value [mm]	Out of tolerance [mm]
5211	0,2309 -0,2309	0,1980	-

Valuation: usable

Operator:

Salvaggio
(Salvaggio)

Date: 22.11.2017

Ref. temp: (20 ± 1) °C. Inspection requirement: The inspections procedure was based on recognised German inspection specifications (VDI/VDE/DGQ/2618). The measuring equipment and standards used are compared regularly with reference standards calibrated by a calibration service accredited by the European Cooperation for Accreditation (EA) and therefore traceable to the national standards of the PTB. Hence the inspection certificate complies with the traceability requirements of DIN EN ISO 9001.