


## CALIBRATION CERTIFICATE

**Customer** : 0155600  
**Identification number** : 102445023  
**Gauge application** : Incoming inspection  
**Standard of thread** : Pipe threads acc. to ISO 7/2:2000 (EN 10226:2005)  
**Type of gauge** : Parallel thread ring gauge for Taper threads  
**Thread designation** : R 3/8  
**Pitch** : 1,3370 mm  
**Measuring method** : Inspection with checking or setting gauge

### Gauge nominal values

**Major diameter min.** : 16,6620 mm  
**Pitch diameter Nennwert** : 15,8060 mm  
**Minor diameter max.** : 14,9500 mm  
**Minor diameter min.** : 14,9220 mm  
**Distance max. to gauge plane** : 0,1337 mm  
**Distance min. to gauge plane** : -0,1337 mm



### Measuring values Pitch diameter with taper check plug gauge

Testing Device ID	Nominal value [mm]	Actual value [mm]	Out of tolerance [mm]
4330	0,1337 -0,1337	0,0030	-

**Valuation: usable**

Operator:  Date: 28.03.2018  
 (Salvaggio)

**Ref.temp:** (20 ± 1) °C. **Inspection requirement:** The inspections procedure was based on recognised German inspection specifications (VDI/VDE/DGQ/2618). The measuring equipment and standards used are compared regularly with reference standards calibrated by a calibration service accredited by the European Cooperation for Accreditation (EA) and therefore traceable to the national standards of the PTB. Hence the inspection certificate complies with the traceability requirements of DIN EN ISO 9001.