



CALIBRATION CERTIFICATE

Customer : 0155600
Identification number : 110296001
Gauge application : Incoming inspection
Standard of thread : Trapezoidal threads acc. to ISO 2901/2903 (DIN 103-9:1985)
Type of gauge : GO thread ring gauge (solid)
Thread designation : Tr 24x5-7e LH
Pitch : 5,0000 mm
Measuring method : "Three balls" method (using T-probe)
Used ball diameter : 2,5503 mm
Constant of T-shaped probe : 16,0805 mm
Measurement traceability : Mahr 828 ID 381
Measurement traceability : Gauge block set ID 3997 57308 D-K-15190-01-00 2015-06

Gauge nominal values


Major diameter min. : 24,5000 mm
Pitch diameter max. : 21,3935 mm
Pitch diameter min. : 21,3605 mm
Minor diameter max. : 19,0165 mm
Minor diameter min. : 18,9835 mm

Measuring values Pitch diameter

| Plane | Axial section | PitchØ [mm] | Out of tolerance [µm] |
|-------|---------------|-------------|-----------------------|
| 2 | 0 Degree | 21,3697 | - |

Valuation: usable

Operator:


 (Salvaggio)

Date: 20.01.2022

Uncertainty of measurement: $U = 2.5 \mu\text{m} + 10 \cdot 10^{-6} \cdot d$. The uncertainty stated is the expanded uncertainty of measurement obtained by multiplying the standard uncertainty by the coverage factor $k = 2$. They were established according to DAkkS-DKD-3. The value of the measured variable is within the assigned value range with 95 % probability.
Ref.temp: $(20 \pm 1) ^\circ\text{C}$. **Inspection requirement:** The inspections procedure was based on recognised German inspection specifications (VDI/VDE/DGQ/2618). The measuring equipment and standards used are compared regularly with reference standards calibrated by a calibration service accredited by the European Cooperation for Accreditation (EA) and therefore traceable to the national standards of the PTB. Hence the inspection certificate complies with the traceability requirements of DIN EN ISO 9001.