

## CALIBRATION CERTIFICATE

**Customer** : 0155600  
**Identification number** : 126883002  
**Gauge application** : Incoming inspection  
**Standard of thread** : Thread gauges according to factory standards  
**Type of gauge** : GO thread plug with NO GO plain plug for minor diameter  
**Thread designation** : Pg 29  
**Pitch** : 1,5880 mm  
**Measuring method** : Three wires method  
**Measurement traceability** : Mahr 828 Nr.22  
**Measurement traceability** : single gauge ID 316275 014252 D-K-15048-01-00 2019-03

| Gauge nominal values      | Go side      | Not Go side |
|---------------------------|--------------|-------------|
| Major diameter max.       | : 37,0400 mm | 35,7500 mm  |
| Major diameter min.       | : 37,0000 mm | 35,7100 mm  |
| Pitch diameter max.       | : 36,2800 mm |             |
| Pitch diameter min.       | : 36,2400 mm |             |
| Kerndurchmesser Höchstmaß | : 35,5200 mm |             |

### Measuring values Pitch diameter - Go side

| Plane | Axial section | PitchØ [mm] | Out of tolerance [µm] |
|-------|---------------|-------------|-----------------------|
| 2     | 0 Degree      | 36,2666     | -                     |

### Measuring values - Not go side

| Plane | Axial section | MajorØ [mm] | Out of tolerance [µm] |
|-------|---------------|-------------|-----------------------|
| 1     | 0 Degree      | 35,7351     | -                     |

Valuation: usable

Operator:

(Ammann)

Date: 10.01.2020

**Uncertainty of measurement:**  $U = 2.5 \mu\text{m} + 10 \cdot 10^{-6} \cdot d$ . The uncertainty stated is the expanded uncertainty of measurement obtained by multiplying the standard uncertainty by the coverage factor  $k = 2$ . They were established according to DAkkS-DKD-3. The value of the measured variable is within the assigned value range with 95 % probability.  
**Ref.temp:** (20 ± 1) °C. **Inspection requirement:** The inspections procedure was based on recognised German inspection specifications (VDI/VDE/DGQ/2618). The measuring equipment and standards used are compared regularly with reference standards calibrated by a calibration service accredited by the European Cooperation for Accreditation (EA) and therefore traceable to the national standards of the PTB. Hence the inspection certificate complies with the traceability requirements of DIN EN ISO 9001.