




CALIBRATION CERTIFICATE

Customer : 0155600
 Identification number : 127719001
 Gauge application : Incoming inspection
 Type of gauge : ring limit gauge
 Thread designation: : DIN 158-M 6 keg.

Method of measuring	Basic pitch Ø (mm)	Toleranc (µm)	effective Ø (mm)
with checking plug gauge	5,275	±4	5,278
Meas.plane	2,000		

Valuation: usable

Remarks: **Effektive diameter check of gauge taper thread plug limit gauge**

Operator: 
 (Salvaggio)

Inspection date: 13.02.2020

Ref.temp: (20 ± 1) °C. Inspection requirement: The inspections procedure was based on recognised German inspection specifications (VDI/VDE/DGQ/2618). The measuring equipment and standards used are compared regularly with reference standards calibrated by a calibration service accredited by the European Cooperation for Accreditation (EA) and therefore traceable to the national standards of the PTB. Hence the inspection certificate complies with the traceability requirements of DIN EN ISO 9001.