



## CALIBRATION CERTIFICATE

**Customer** : 0155600  
**Identification number** : 127835005  
**Gauge application** : Incoming inspection  
**Standard of thread** : Cyl. pipe threads "NPSM" acc. to ANSI/ASME B1.20.1-2013  
**Type of gauge** : GO/NOT GO thread plug gauge  
**Thread designation** : 3/8 - 18 NPSM-2B  
**Pitch** : 1,4110 mm  
**Measuring method** : Three wires method  
**Wire diameter** : 0,8662 mm  
**second wire diameter** : 0,8662 mm  
**Measuring force** : 9,00 N  
**Measurement traceability** : Mahr 828 Nr.22  
**Measurement traceability** : Gauge block set ID 3997 57308 D-K-15190-01-00 2015-06

Gauge nominal values	Go side	Not Go side
Major diameter max.	: 16,8549 mm	16,6868 mm
Major diameter min.	: 16,8422 mm	16,6741 mm
Pitch diameter max.	: 15,9334 mm	16,0757 mm
Pitch diameter min.	: 15,9258 mm	16,0680 mm
Minor diameter max.	: 15,1115 mm	15,1115 mm


### Measuring values Pitch diameter - Go side

Plane	Axial section	PitchØ [mm]	Out of tolerance [µm]
2	0 Degree	15,9294	-

### Measuring values Pitch diameter - Not go side

Plane	Axial section	PitchØ [mm]	Out of tolerance [µm]
2	0 Degree	16,0682	-

**Valuation: usable**

Operator:   
(Sahan)

Date: 09.04.2019

**Uncertainty of measurement:**  $U = 2.5 \mu\text{m} + 10 \cdot 10^{-6} \cdot d$ . The uncertainty stated is the expanded uncertainty of measurement obtained by multiplying the standard uncertainty by the coverage factor  $k = 2$ . They were established according to DAkkS-DKD-3. The value of the measured variable is within the assigned value range with 95 % probability.  
**Ref.temp:** (20 ± 1) °C. **Inspection requirement:** The inspections procedure was based on recognised German inspection specifications (VDI/VDE/DGQ/2618). The measuring equipment and standards used are compared regularly with reference standards calibrated by a calibration service accredited by the European Cooperation for Accreditation (EA) and therefore traceable to the national standards of the PTB. Hence the inspection certificate complies with the traceability requirements of DIN EN ISO 9001.