

CALIBRATION CERTIFICATE

Customer : 0155600 Identification number : 130198119

Gauge application : Incoming inspection

Standard of thread : Metric ISO threads acc. to ISO 965:2013/ISO 1502:1996

Type of gauge : GO/NOT GO thread plug gauge

Thread designation : M 4x0.7-6H Pitch : 0,7000 mm

Measuring method : Three wires method

Wire diameter : 0,4040 mm second wire diameter : 0,4040 mm Measuring force : 4,10 N

Measurement traceability : Mahr 828 Nr.31

Measurement traceability : Gauge block set ID 3997 57308 D-K-15190-01-00 2015-06

Gauge nominal values Go side Not Go side Major diameter max. 4,0150 mm 3,8165 mm Major diameter min. 3,9970 mm 3,7985 mm Pitch diameter max. 3,5555 mm 3,6720 mm Pitch diameter min. 3,5465 mm 3,6630 mm Minor diameter max. 3,1410 mm 3,1410 mm

Measuring values Pitch diameter - Go side

(Eisele)

| Plane | Axial section | PitchØ [mm] | Out of tolerance [µm] |
|-------|---------------|-------------|-----------------------|
| 2 | 0 Degree | 3,5512 | - |

Measuring values Pitch diameter - Not go side

| Plane | Axial section | PitchØ [mm] | Out of tolerance [µm] |
|-------|---------------|-------------|-----------------------|
| 2 | 0 Degree | 3,6690 | - |

Valuation: usable

Operator:

Date: 19.11.2020

Uncertainty of measurement: U = 2.5 µm + 10 • 10⁻⁶ • d. The uncertainty stated is the expanded uncertainty of measurement obtained by multiplying the standard uncertainty by the coverage factor k = 2. They were established according to DAkkS-DKD-3. The value of the measured variable is within the assigned value range with 95 % probability.

Ref.temp: (20 ± 1) °C. Inspection requirement: The inspections procedure was based on recognised German inspection specifications (VDI/VDE/DGQ/2618). The measuring equipment and standards used are compared regularly with reference standards calibrated by a calibration service accredited by the Eropean Cooperation for Accreditation (EA) and therefore traceable to the national standards of the PTB. Hence the inspection certificate complies with the tracebility requirements of DIN EN ISO 9001.