



## CALIBRATION CERTIFICATE

**Customer** : 0155600  
**Identification number** : 131204004  
**Gauge application** : Incoming inspection  
**Standard of thread** : Pipe threads acc. to ANSI/ASME B1.20.1-1983 (NPT)  
**Type of gauge** : Taper full form thread ring gauge  
**Thread designation:** : 1/16-27 NPT (4,064)  
**Measuring method** : with checking plug  
**Identification number** : 6706

<i>Method of measuring</i>	<i>Basic pitch Ø (mm)</i>	<i>Toleranc (µm)</i>	<i>effective Ø (mm)</i>
with checking plug gauge	7,142	±2,5	7,1396
Meas.plane	4,064		

**Valuation: usable**

**Remarks: Effektive diameter check of gauge taper thread plug limit gauge**

Operator: .....  
(Salvaggio)

Inspection date: 29.04.2019

**Ref.temp:** (20 ± 1) °C. **Inspection requirement:** The inspections procedure was based on recognised German inspection specifications (VDI/VDE/DGQ/2618). The measuring equipment and standards used are compared regularly with reference standards calibrated by a calibration service accredited by the European Cooperation for Accreditation (EA) and therefore traceable to the national standards of the PTB. Hence the inspection certificate complies with the traceability requirements of DIN EN ISO 9001.