



Talstraße 19
D-72461 Albstadt

Tel. ++49 (0) 74 32-90 87-0
Fax ++49 (0) 74 32-90 87-60

contact@johs-boss.de
www.johs-boss.de

CALIBRATION CERTIFICATE

Customer : 0155600
 Identification number : 266643
 Gauge application : Incoming inspection
 Standard of thread : Metrical threads according DIN ISO 1502 (DIN 13)
 Type of gauge : solid GO thread ring gauge
 Thread designation: : M 21x0.75-6g
 Pitch: : 0,7500 mm
 Measuring method: : "Three balls" method (using T-probe)
 Used ball diameter: : 0,4555 mm
 Constant of T-shaped probe : 2,4113 mm

Gauge nominal values

External diameter min. value : 21,0410 mm
 Effective diameter - min. value : 20,4820 mm
 Effective diameter - max. value : 20,4960 mm
 Minor diameter - min. value : 20,1590 mm
 Minor diameter - max. value : 20,1730 mm

Measuring values of Effective diameter

Meas. plane No.	Meas. value No.	Effective diameter [mm]	Out of tolerance [μm]
1	1	20,4889	-

Valuation: in tolerance

Operator:
(Salvaggio)

Inspection date: 10.07.2008

measuring accuracy: U95 = 2.5 μm ; **Ref.temp.** [deg C°]: 20; **Measuring machine:** MAHR 828 Nr.19372; MAHR 828 Nr. 38013; Zeiss ULM 01-600C Nr.6908; IAC Master Scanner MSXP 6025 Nr.101064; IAC Master Scanner MSXP 10060 Nr.100971; **reference standard:** Setting ring gauge \varnothing 20 342-DKD-K-14201-05-07; Parallel gauge block: 10mm 5328-DKD-K-17301-2005-07; **inspection requirement:** The inspections procedure was based on recognised German inspection specifications (VDI/VDE/DGQ/2618). The measuring equipment and standards used are compared regularly with reference standards calibrated by a calibration service accredited by the European Co-operation for Accreditation (EA) and therefore traceable to the national standards of the PTB*. Hence the inspection certificate complies with the traceability requirements of DIN EN ISO 9000:2000, section 7.6.

*Physikalisch Technische Bundesanstalt (Federal Physical and Technical Institute)